



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

HOWELL LABORATORIES INC.
188 Harrison Road
Bridgton, ME 04009
Peter Heang Phone: 207 647 3327
pheang@howelllabs.com

CALIBRATION

Valid To: November 30, 2026

Certificate Number: 4823.02

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations^{1,3}:

I. Thermodynamics

Parameter/Equipment	Range	CMC ² (±)	Comments
Humidity – Dew Point Sensors	-100 °C ⁴ -90 °C -80 °C (-60 to 20) °C	1.2 °C 1.2 °C 0.74 °C 0.40 °C	Chilled mirror dew point meter

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ This scope meets A2LA's *PI12 Flexible Scope Policy*.

⁴ This temperature is not directly traceable; it is considered consensus traceability.



Accredited Laboratory

A2LA has accredited

HOWELL LABORATORIES INC.

Bridgton, ME

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 23rd day of October 2024.

A blue ink signature of Trace McInturff, written over a horizontal line.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 4823.02
Valid to November 30, 2026

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.